

Daily Care and Drying

1. Open the beaker and remove the inside container.



2. Remove the drying capsule from the packaging and place it in the bottom of the beaker with the orange side facing up. Drying capsules should be replaced every 2-3 months or when the orange colour has faded.



3. Reinstall the inside container.



4. Remove any receivers and/or audioshoes from the hearing aids. Place the hearing aids, receivers, and audioshoes in the beaker with the battery door open.

